

93-7585 Meeting 10/15/98

Attendees

Jim Stafford
Stacy Wade
John Skuba
Luke Hinkle
Bob Maxwell
Jim Poulin
Jack Gillespie

Discussed current status of special at Applied.

Will provide Jim Poulin with costs and drawings of the Pirani sensor so they can get a better handle on the project.

Calibration of the switch needs to be done after welding; however, it is not a difficult procedure and can be done at the service centers. Simply need system with closed loop pressure control.

Did not get a good feeling anyone was ready to pick up and run with this ball.

jws